



Immunosuppressant (IPT) IP0402 - Allgemeines Krankenhaus Wien Individual Report

Round: 417

Issue Number 1

Issued 08 January 2019



Science
for a safer world



Scheme: Immunosuppressant (IPT)

Round: 417

CICTAC - Ciclosporin and/or Tacrolimus

Analyte	Analyst	Method	Result	Units	z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
A - Ciclosporin	Lab Result	HPLC-MS	821.00	µg/L	0.61	785.00	6.58	59.32	N/A	129	785.00	787.54	59.320	73.095
B - Ciclosporin	Lab Result	HPLC-MS	158.00	µg/L	0.46	152.74	1.27	11.493	N/A	129	152.74	152.94	11.493	13.088
C - Ciclosporin	Lab Result	HPLC-MS	145.00	µg/L	0.07	144.00	1.49	13.347	N/A	129	144.00	144.65	13.347	14.450
A - Tacrolimus	Lab Result	HPLC-MS	7.00	µg/L	-0.10	7.04	0.04	0.408	N/A	140	7.04	7.02	0.408	0.592
B - Tacrolimus	Lab Result	HPLC-MS	24.60	µg/L	0.22	24.30	0.15	1.364	N/A	140	24.30	24.45	1.364	1.785
C - Tacrolimus	Lab Result	HPLC-MS	4.20	µg/L	0.00	4.20	0.03	0.297	N/A	140	4.20	4.20	0.297	0.377

SIR- Sirolimus

Analyte	Analyst	Method	Result	Units	z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
A - Sirolimus	Lab Result	HPLC-MS	10.20	µg/L	0.15	10.00	0.16	1.335	N/A	116	10.00	9.89	1.335	1.351
B - Sirolimus	Lab Result	HPLC-MS	4.60	µg/L	-0.52	4.95	0.08	0.667	N/A	116	4.95	4.92	0.667	0.759
C - Sirolimus	Lab Result	HPLC-MS	19.90	µg/L	-0.36	20.80	0.29	2.521	N/A	116	20.80	20.88	2.521	2.723

** Please note, participant performance for this analyte has been assessed using a z' score, rather than a z score, in order to account for the measurement uncertainty of the assigned value which is not negligible when compared to the SDPA.