

Centre 145

PT Distribution 363

Samples dispatched on: 3rd June 2014
Results deadline: 27th June 2014


EINGEGANGEN
02 Juli 2014
TQM




Bias plot: Distribution 360 - 363

Method: AC/MIA – on Dimension platform


Method: AC/MIA – on Dimension platform




Group Code:	AC/MIA – on Dimension platform	900.0
Target concentration (ng/mL)		900.0
Method mean (ng/mL)		901.1
SD		107.3
CV		11.9%
Your result (ng/mL)		773.0
Inaccuracy		-14.2%
Z Value		-1.2
Accepted		Yes
Total number of results		62
Number of accepted results		61



Group Code:	AC/MIA – on Dimension platform	50.0
Target concentration (ng/mL)		50.0
Method mean (ng/mL)		47.9
SD		8.9
CV		18.6%
Your result (ng/mL)		34.0
Inaccuracy		-29.0%
Z Value		-1.6
Accepted		Yes
Total number of results		62
Number of accepted results		59



Group Code:	AC/MIA – on Dimension platform	250.0
Target concentration (ng/mL)		250.0
Method mean (ng/mL)		249.6
SD		18.9
CV		7.6%
Your result (ng/mL)		229.0
Inaccuracy		-8.3%
Z Value		-1.1
Accepted		Yes
Total number of results		62
Number of accepted results		61



PT Code	360A	360B	360C	361A	361B	361C	362A	362B	362C	363A	363B	363C
Bias	-4.4%	3.4%	-3.4%	-12.7%	-12.2%	-0.9%	-8.1%	-6.2%	-14.2%	-29.0%	-8.3%	-29.0%

Target value	1000.0	100.0	Pool	30.0	Pool	100.0	450.0	Pool	60.0	Pool	900.0	50.0	250.0
Method Mean	973.5	99.6	126.3	34.1	138.6	103.7	467.4	162.1	67.2	901.1	47.9	249.6	
Your value	931.0	103.0	122.0	<30	121.0	91.0	463.0	149.0	63.0	773.0	34.0	229.0	
Accepted	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	